

# Sneak Peek

*Design Control and Risk Management  
For Combination Products*



**James Pink**

Health Sciences and Medical  
Devices Consultant



**SYMMETRIC**

# Design plan - inputs

01

	<b>Pen Injector</b>	<b>Transdermal Patch</b>	<b>Metred dose inhaler</b>
<b>Regulatory requirements</b>	US Product code NSC	US Product code QIX	US Product code QKS
<b>FDA / EU / Guidance</b>	ISO 7864 ISO 9626 ISO 23908		
<b>Product Standards</b> <ul style="list-style-type: none"><li>International / EU / Consensus (US) / 510K data</li><li>Industry standards</li></ul>	SO 11608-4	?	ISO 20072 ISO 27427
<b>Process / Material Standards</b>	ISO 20069 USP39_NF34<85> ISO 10993-1, 5, 10 ISO 11608-2 ASTM F1929098 ISTA 3A	ISO 20069 ISO 11135	ISO 20069
	Risk analysis	Risk analysis	Risk analysis

# Design Input

02




Figure 3 : Determining sources of requirements and translating to approved design input

# Requirements - best practice

03

Objective	Example
Provide the source of the requirement	Standard number and clause, marketing, risk management activity
Be specific	Component / Phase / User step
Use imperatives	Shall or Must
Ensure the requirement is testable	Describe an expected outcome i.e. compliance to a standard test method.
Structure the requirement to be consistent	ID, Type, Component, Phase, Output etc
Don't be subjective	Avoid words like "normal, strengthen, enhance, user-friendly". Opt for precise and measurable terms.

# Risk management in combination products

04

- **Failure to actuate, use-related failures**

The purpose of this letter is to inform you that the administration of EpiPen® 0.3 mg and EpiPen Jr® 0.15 mg Auto-Injectors and the authorized generic versions (referenced collectively as EpiPen in the remainder of this letter) may be delayed or prevented during an emergency due to:

1. Device failure from spontaneous activation caused by using a sideways force to remove the blue safety release.
2. Device failure from inadvertent or spontaneous activation due to a raised blue safety release
3. Difficulty removing the device from the carrier tube.
4. Certain identified use errors.



# Registration

05

## Sneak Peek

*Design Control and Risk Management  
For Combination Product*

**CLICK  
HERE TO**




---

Mliekarenská 9, 821 09  
Bratislava, Slovak Republic  
ID: 47 068 124  
VAT no: SK2023741973  
Office: +421 948 262 346

🛡 100% Secure payments

Your details are protected and safe with us. Taxes calculated at the checkout.





SYMMETRIC

# Your Partner for Pharma, Biotech and Medical Device Training

We train clients from over 500 companies annually. Enter our global educational platform for pharmaceutical professionals.

[VIEW ALL TRAINING COURSES](#)